


## PeriView FLEX TBNA Needle Quick Reference Guide

Below are steps for PeriView FLEX single-use aspiration needle (TBNA) that is designed to safely obtain samples from the intrapulmonary region.

**CAUTION:** This guide is only a summary of the PeriView FLEX TBNA Needle Instructions for Use. For detailed operating instructions, be sure to follow the PeriView FLEX TBNA Needle instruction manual included in the device package.

### Product Overview



PeriView FLEX



PeriView FLEX TBNA Needle with and without removable stylet



Optional suction syringe (not included)



Single-use adaptor biopsy valve (not included)

### PeriView FLEX Steps for Use



Remove device by pulling out handle first from the sterile pouch and card.



Inspect the device for appearance and operation. Ensure the slider is fully in the retracted position. Prepare guide sheath and devices.

# PeriView FLEX TBNA Needle Quick Reference Guide

**CAUTION:** This guide is only a summary of the PeriView FLEX TBNA Needle Instructions for Use. For detailed operating instructions, be sure to follow the PeriView FLEX TBNA Needle instruction manual included in the device package.

## PeriView FLEX Steps for Use (continued)



Once the desired location has been reached with the endoscope, slowly insert the distal end of the device into the guide sheath or biopsy valve on the endoscope's instrument channel port.



If additional flexibility is required to access target, partially or fully retract the stylet as needed.



Pierce the target tissue with the needle by pushing the needle slider to extend the needle to the desired length (maximum 20 mm).



Remove the stylet.



If suction is preferred, attach the prepared syringe.



Move the needle slider up and down several times while the needle is in the target area to obtain a sample.



If utilizing suction, release suction and remove the syringe.



Fully retract the needle into the sheath by pulling the needle slider proximally until it clicks.



Withdraw the device from the endoscope.


# PeriView FLEX TBNA Needle Quick Reference Guide

**CAUTION:** This guide is only a summary of the PeriView FLEX TBNA Needle Instructions for Use. For detailed operating instructions, be sure to follow the PeriView FLEX TBNA Needle instruction manual included in the device package.

## Slide Preparation




Push down on the needle slider to extend the needle. Then connect 5 to 20cc air filled syringe to the aspiration port.



Gently push the syringe plunger to expel a small drop of specimen on the slide. Dispense the remaining specimen into the cell block container. Repeat this step as needed until entire specimen is ejected.



**Alternative:** If the specimen cannot be expelled with air alone, remove the syringe from the aspiration port and insert the stylet into the aspiration port to push out the specimen.



Expel a small drop of specimen on the slide. Dispense the remaining specimen into the cell block container.

## Reinserting the Device for Additional Sampling



Flush the inside of the needle with a syringe filled with 3-5 cc of sterile saline followed by 20 cc of air.

Olympus is a registered trademark of Olympus Corporation, Olympus America Inc., and/or their affiliates.