

OLYMPUS

EBUS 180 (BF-UC180F)

CAPS ARE MH-553

TAKE THE x2 CAPS OFF FOR STERILISATION



EBUS 190 (BF-UC190F)

MAJ-2295

LEAVE THE X1 CAP ON FOR STERILISATION

Please refer to the Olympus Manufacturer Reprocessing Manuals for further information.

OLYMPUS AUSTRALIA PTY LTD
3 Acacia Place, Notting Hill, VIC 3168
PO Box 985 Mt Waverley, VIC 3149
Phone: 1300 132 992 Fax: 03 9543 1350
Web: www.olympusaustralia.com.au
ABN 90 078 493 295

OLYMPUS NEW ZEALAND LIMITED
28 Corinthian Drive, Albany, Auckland NZ 0632
Phone: 0508 6596787 Fax: 09 836 3386
Web: www.olympus.co.nz
Companies No: 910603

